

## CERTIFICATE OF ANALYSIS

<b>Product Name</b>	WATER For HPLC	<b>Product Code</b>	HPW01-1000
<b>Batch Number</b>	W01-210801	<b>Molecular Formula</b>	H <sub>2</sub> O
<b>Mfg. Month</b>	August - 2021	<b>Best Before</b>	July - 2022
<b>CAS Number</b>	7732-18-5	<b>Batch Lot</b>	2000 Liter
<b>Packing Type</b>	Amber Pet Bottle	<b>Pack Size</b>	1 Liter

### Result Of Analysis

Parameters	Specification Limits	Results
Appearance	Clear Colorless Liquid	Passes Test
Conductivity at 30°C, max	0.25 µS/cm	Passes Test
Resistivity at 30°C, max	4.0 MΩ·cm	Passes Test
pH Value @ 25°C	5 - 7	Passes Test
TOC Count	≤ 200 ppb	Passes Test
Evaporation Residue, max	0.0005%	Passes Test
Sodium, max	0.000001%	Passes Test
Chloride, max	0.000001%	Passes Test
Silica, max	0.00005%	Passes Test
Sulphate, max	0.00001%	Passes Test
Suitability For HPLC	Complies	Passes Test
<b>UV Absorption in 1cm cell v/s H<sub>2</sub>O</b>		
UV Absorption @210nm	0.01 %	Passes Test
UV Absorption @254nm	0.005 %	Passes Test

\* Disclaimer – All Results shared of the parameters aforementioned are observed at the time of filling.

### STORAGE

Store in original container protected from direct sunlight in a dry, cool and well-ventilated area.

Quality Assurance Head

Company Seal